

시험 성적서

(TEST REPORT)



한국산업기술시험원
Korea Testing Laboratory

성적서 번호 : 19-084077-01-1
Report No.

페이지 (1) / (총 2)
Page of Pages



1. 의뢰자 (Client)

기관명 (Name) : 게이트비전 주식회사

주소 (Address) : 경기도 고양시 일산동구 일산로 111 (백석동, 게이트비전빌딩)

의뢰일자 (Date of Receipt) : 2019. 12. 18.

2. 시험성적서의 용도 (Use of Report) : 품질관리용

3. 시험대상품목/물질/시료명 (Test Sample)

제품명 (Description) : 공기살균기

제조회사 (Manufacturer) : NOVAERUS

모델명 (Model Name) : NV330

제조번호 (Serial Number) : ***

기타 (Remark) : ***

4. 시험기간 (Date of Test) : 2019년 12월 19일 ~ 2019년 12월 24일

5. 시험장소 (Location of Test) :

☒ KTL 고정시험실 (주소 : 서울특별시 구로구 디지털로26길 87 (구로동))

☐ 현장시험

6. 시험규격/방법 (Test Standard/Method) : KOUVA AS 02 : 2017

7. 시험결과 (Test Results) : 후면참조

- 비고(Note) :
1. 이 성적서는 의뢰자가 제출한 시료에 한하며, 법적 및 기타분쟁의 근거 등으로의 사용을 금합니다.
 2. 이 성적서는 원본만 유효하며, 임의로 재가공된 사본 및 전자인쇄본 등은 유효하지 않습니다.
(‘원본’이란 KTL에서 정해진 절차에 따라 보안성을 포함시켜 제공하는 모든 성적서를 의미합니다.)
 3. 아래의 2D바코드를 스캔하여 성적서의 원본내용 확인이 가능하며, KTL 보관 원본과의 동일성은 고객전용홈페이지(customer.ktl.re.kr)의 "성적서 원본확인"창에서 비교가능 합니다.

확인 Affirmation	작성자(Tested by)	기술책임자(Technical Manager)
	성명(Name): 박서현 (Signature)	성명(Name): 전용우 (Signature)


2020. 01. 13.

한국산업기술시험원



서울특별시 구로구 디지털로26길 87 (구로동) (87, Digital-ro 26-gil, Guro-gu, Seoul, KOREA) Tel.02-860-1139 Fax. 02-860-1689

FP204-01-06



※ 위 마크는 추후 전자확인증 대조 프로그램에서 원본대조시 사용되는 2D코드입니다.

시험결과 (Test Results)

1. 시험결과

품명	기준 및 결과	기준	부유바이러스 저감율
공기살균기		60 % 이상	82.0 %

2. 시험품


(1) 제 조 사 : NOVAERUS

(2) 모 델 : NV330

3. 시험방법 및 조건


시험방법	바이러스	온도	습도	시험챔버	시험시간	풍량레벨
KOUVA AS 02:2017, 공기 정화장치	Phi-X174 (ATCC 13706-B1)	(23 ± 2) °C	(50 ± 5) % R.H.	60 m ³	30 분	단일모드

4. 시험품 사진



이 하 여 백

FP204-02-02



※ 위 마크는 추후 전자확인증 대조 프로그램에서 원본대조시 사용되는 2D코드입니다.

TEST REPORT



한국산업기술시험원
Korea Testing Laboratory

Report No. : 19-084077-02-1

Page of Pages : (1) / (2)



1. Client

Name : GATEVISION

Address : GATEVISION Bldg. 111, ilsan-ro, ilsandong-gu, Goyang-si, Gyeonggi-do, Korea

Date of Receipt : 2019. 12. 18.

2. Use of Report : For quality control

3. Test Sample

Description : Air sterilizer

Manufacturer : NOVAERUS (Ireland) Ltd.

Model Name : NV330

Serial Number : ***

Remark : ***

4. Date of Test : 2019. 12. 19. ~ 2019. 12. 24.

5. Location of Test :


☒ KTL Permanent Test (Address : 87, Digital-ro 26-gil, Guro-gu, Seoul, KOREA)

☐ On Site Testing

6. Test Standard/Method : KOUVA AS 02 : 2017

7. Test Results : Refer to attached results

- Note : 1. This report is limited to samples submitted by the applicant and is prohibited from being used for legal or other reasons of dispute.
2. This document is valid only in its original document, and any reproduced copies and electronic copies are not valid. ("Original" means all the reports provided by the KTL including the security procedures.)
3. You can check the contents of the report by scanning the 2D Barcode below. The identity of original reports can be checked in the "Confirm original report" window of the customer's homepage (customer.ktl.re.kr).


Affirmation	Tested by	 (Signature)	Technical Manager	 (Signature)
	Name : Park seo-hyun		Name : JEON, Yong-Woo	

2020. 01. 13.

Korea Testing Laboratory

87, Digital-ro 26-gil, Guro-gu, Seoul, KOREA Tel.+82-2-860-1139 Fax. +82-2-860-1689

FP204-03-06



※ 위 마크는 추후 전자확인증 대조 프로그램에서 원본대조시 사용되는 2D코드입니다.

Test Results

1. Test Results

	Reference value	Results
Air sterilizer	over 60 %	82.0 %


2. Test sample

- (1) Manufacturer : NOVAERUS (Ireland) Ltd.
- (2) Model name : NV330
- (3) Serial number : ***

3. Test Method and Conditions

Test method	Conditions					
	Virus	Temp.	Humidity	Test time	Test Chamber	Air flow
KOUVA AS 02 : 2017	Phi-X174 (ATCC 13706-B1)	(23 ± 2) °C	(50 ± 5) % R.H.	30 minute	60 m ³	Single mode

4. Unit Photograph



- End of document -

FP204-04-02



※ 위 마크는 추후 전자확인증 대조 프로그램에서 원본대조시 사용되는 2D코드입니다.